

FSP GROUP INC  
 MR A CHENG  
 22 JIANGUO E RD  
 TAOYUAN CITY TAIWAN

**RE: Project Number(s) - 03CA10362**

Your most recent Certification is shown below. You may also view this information, or a portion of this information (depending on the product category), on UL's Online Certifications Directory at [www.ul.com/database](http://www.ul.com/database). Please review the text and contact the Conformity Assessment Services staff member who handled your project if revisions are required. For instructions on placing an order for this information in a 3 x 5-inch format, you may refer to the enclosed order form for UL Card Service.

**QQHM2** April 14, 2003  
**Power Supplies, Medical and Dental - Component**

**FSP GROUP INC**  
**22 JIANGUO E RD, TAOYUAN CITY TAIWAN**

**E211696**

Model No.	Rated Input		SC	Output			OC	SP	EP	FC	GC
	V	Hz		Max V	Max A	Max VA					
FSP200-60MD(a)	100-120/ 200/240	60/50	0	+5.0	18	—	9	2601	20B	0	1
				+12.0	6	—					
				+3.3	14	—					
				-5.0	0.3	—					
				-12.0	0.2	—					
				+5sb	2.0	—					
FSP300-60MD(a)	100-120/ 200/240	60/50	0	+5.0	30	—	9	2601	20B	0	1
				+12.0	10	—					
				+3.3	14	—					
				-5.0	0.3	—					
				-12.0	0.8	—					
				+5sb	2	—					
FSP200-50NIV(a)	115/230	60/50	0	+3.3	16.7	—	9	2601	20B	0	1
				+5	16	—					
				+12	10	—					
				-12	0.3	—					
				+5 V sb	2	—					

(a)Output values are rated, not maximum.  
 Marking: Company name and model designation.

**See General Information Preceding These Recognitions**  
 For use only in equipment where the acceptability of the combination is determined by Underwriters Laboratories Inc.